Model CB 260 | CO2 incubators with hot air sterilization and heat sterilizable CO2 sensor

The powerful and versatile CB CO₂ incubator is suitable for cell-based assays in drug and medical research. This BINDER incubator is even suitable for complex cultivation experiments or individual environments with hypoxic conditions thanks to its comprehensive assortment of options and accessories.

BENEFITS

- Safe cell cultures receive maximum protection against contamination
- Reliable constant well-being conditions for the cells
- Smart simple routine cleaning and convenient operation
- Economical efficient operation without consumables





Model 260



IMPORTANT FEATURES

- Optimized double-pan humidification system with condensation protection
- Tried-and-tested anti-contamination concept with 180°C hot air sterilization
- Hot-air sterilizable CO2 sensor with single-beam infrared technology
- Seamless inner chamber made of stainless steel with flanges as shelf support
 system
- Fanless interior with Venturi CO_2 gas mixing nozzle

- Intuitive touchscreen controller
- Internal data recording, can be read out in open format via USB interface
- Fault diagnosis system with visual and acoustic alarm
- Interfaces: Ethernet, USB, zero-voltage alarm contact
- Fail-safe CO2 system for protecting the pH of cell cultures

OPTIONAL EQUIPMENT

- O_2 control with O_2 control ranges: 0.2 to 20 vol.% O_2 or 10 to 90 vol.% O_2
- Multi-sectional glass doors with special shelves for dividing up the space to accommodate different experiments
- Stacking adapter for secure, space-saving installation
- Gas tank changer for connecting two gas tanks to one or two incubators

ORDERING INFORMATION

Interior volume [L]	Power supply - unit fuse	Plug*	Version	Model version	ArtNo.
Model CB 260					
	200 2001/4 50//011 40.04		Standard	CB260-230V	9640-0013
267	200230 V 1- 50/60 Hz -10,0 A	CEE 7/7	with O ₂ control 0.2-20%	CB260-230V-O	9640-0014
	100120 V 1~ 50/60 Hz -16,0 A	NEMA 5-20	Standard	CB260UL-120V	9640-0015

BINDER Best conditions for your success

Interior volume [L]	Power supply - unit fuse	Plug*	Version	Model version	ArtNo.
			with O ₂ control 0.2-20%	CB260UL-120V-0	9640-0016

Data		
Designation	CB260-230V	CB260-230V-O
Article Number	9640-0013	9640-0014
Option model	Standard	with O2 control 0.2-20%
Performance Data Temperature		
Temperature range	+5 $^{\circ}\text{C}$ above ambient temperature to 60 $^{\circ}\text{C}$	+6 $^{\circ}\text{C}$ above ambient temperature to 60 $^{\circ}\text{C}$
Temperature uniformity at 37°C	0.4 ± K	0.4 ± K
Temperature fluctuation at 37°C	0.1 ± K	0.1 ± K
Recovery time after door was opened for 30 s at 37 °C	6 min	6 min
Performance Data Climate	-	-
Humidity range	9095 % RH	9095 % RH
Gas Data		
(0. rango		

Data			
Designation	CB260-230V	CB260-230V-O	CB260UL-120V
Article Number	9640-0013	9640-0014	9640-0015
Option model	Standard	with O2 control 0.2-20%	Standard
Performance Data Temperature			-
Temperature range	+5 $^\circ\text{C}$ above ambient temperature to 60 $^\circ\text{C}$	+6 $^{\circ}\text{C}$ above ambient temperature to 60 $^{\circ}\text{C}$	+5 °C above ambient temper to 60 °C
Temperature uniformity at 37°C	0.4 ± K	0.4 ± K	0.4 ± K
Temperature fluctuation at 37°C	0.1 ± K	0.1 ± K	0.1 ± K
Recovery time after door was opened for 30 s at 37 °C	6 min	6 min	6 min
Performance Data Climate			
Humidity range	9095 % RH	9095 % RH	9095 % RH



CB260UL-120V-0

with O2 control 0.2-20%

+6 $\,^\circ\text{C}$ above ambient temperature

9640-0016

to 60 $^\circ\text{C}$

Temperature uniformity at $37^{\circ}C$	0.4 ± K	0.4 ± K	0.4 ± K	0.4 ± K
Temperature fluctuation at 37°C	0.1 ± K	0.1 ± K	0.1 ± K	0.1 ± K
Recovery time after door was opened for 30 s at 37°C	6 min	6 min	6 min	6 min
Performance Data Climate	-			
Humidity range	9095 % RH	9095 % RH	9095 % RH	9095 % RH
Gas Data	-	-		
CO ₂ range	020 Vol% CO2	020 Vol% CO2	020 Vol% CO2	020 Vol% CO2
CO2 measuring technology	IR	IR	IR	IR
CO2 recovery time after door was opened for 30 s at 5 vol. % CO2	5 min	5 min	5 min	5 min
Standard O2 control range		0.220 Vol% O2		0.220 Vol% O2
O2 control ranges with option: O2 range		1090 Vol% O2		1090 Vol% O2
O2 recovery time after door was opened for 30 s at 5 vol. % O2		15 min		15 min
Electrical data				
Rated Voltage	200230 V	200230 V	100120 V	100120 V
Power frequency	50/60 Hz	50/60 Hz	50/60 Hz	50/60 Hz
Nominal power	1,5 kW	1,5 kW	1,5 kW	1,5 kW
Unit fuse	10,0 A	10,0 A	16,0 A	16,0 A
Phase (Nominal voltage)	1~	1~	1~	1~
Dimensions and weights				
Interior volume	267 L	267 L	267 L	267 L
Net weight of the unit (empty)	125 kg	125 kg	125 kg	125 kg


Data s	sheet	Model	СВ	260
--------	-------	-------	----	-----



Load per rack	10 kg	10 kg	10 kg	10 kg
Permitted load	40 kg	40 kg	40 kg	40 kg
Wall clearance back	100 mm	100 mm	100 mm	100 mm
Wall clearance sidewise	50 mm	50 mm	50 mm	50 mm
Housing dimensions not incl. fitting	s and connections			
Width net	740 mm	740 mm	740 mm	740 mm
Height net	1,020 mm	1,020 mm	1,020 mm	1,020 mm
Depth net	785 mm	785 mm	785 mm	785 mm
Internal Dimensions				
Interior width	620 mm	620 mm	620 mm	620 mm
Interior height	750 mm	750 mm	750 mm	750 mm
Interior depth	575 mm	575 mm	575 mm	575 mm
Inner doors	1	1	1	1
Unit doors	1	1	1	1
Environment-specific data		·	-	
Sound-pressure level	41 dB(A)	41 dB(A)	41 dB(A)	41 dB(A)
Energy consumption at 37°C	65 Wh/h	65 Wh/h	65 Wh/h	65 Wh/h
Fixtures	•			
Number of shelves (std./max.)	3/8	3/8	3/8	3/8

All technical data is specified for unloaded units with standard equipment at an ambient temperature of ± 22 °C ± 3 °C and a power supply voltage fluctuation of $\pm 10\%$. The temperature data is determined to BINDER factory standard following DIN 12880, observing the recommended wall clearances of 10 % of the height, width, and depth of the inner chamber. All indications are average values, typical for units produced in series. We reserve the right to change technical specifications at any time.

DIMENSIONS Incl. fittings and connections [mm]



OPTIONS

Designation	Description	CB 260	*	ArtNo.
Access port with silicone plug	back			

Data sheet Model CB 260

BINDER SS

Best	conditions	for	your	succes

Designation	Description	CB 260	*	ArtNo.
	30 mm	•	01	8612-0025
	left			
	30 mm	•	01	8612-0026
	right			
	30 mm	•	01	8612-0027
Analog output 4-20 mA	for temperature and CO_2 values (outputs not adjustable)	•	02	8612-0022
	for temperature, measurement in center of chamber at specified temperature	•	-	8012-1132
Calibration certificate, temperature	temperature measurement incl. certificate, 9 measuring points at specified temperature		-	8012-1550
	temperature measurement incl. certificate, 15- 18 measuring points at specified temperature	•	-	8012-1571
	temperature measurement incl. certificate and 27 measuring points at specified temperature	•	-	8012-1592
Calibration certificate, temperature and CO2	for temperature and CO2, temperature measurement in center of chamber, CO2 measurement performed using test gas at 37 $^\circ$ C and 5 $\%$ CO2	•	-	8012-1235
	for versions with O_2 control; a measuring point in center of chamber at 80 Vol% O_2	•		8012-2604
Calibration certificate O2	for versions with O_2 control; a measuring point in center of chamber at 1 Vol% O_2	•	-	8012-2603
ELLROLL Set	modular, expandable roller bottle system for cell cultivation, 6 roller bottles	•	-	8612-0024
Door hinged on the left	Unit door and standard glass inner door with door hinged on the left	•	-	8612-0034
Electric access port	8-pin, for low voltage with LEMO socket (coverable) and LEMO plug (max. 24 V \cdot 2 A)		-	8612-0033
nner door, divided	Cell therapy compartmentalization, consisting of 8 inner doors and four shelf levels with one partition wall each	•	-	8612-0031
)2 control range	alternative to the standard range, for controlling O_2 concentrations between 10 and 90 vol. %, only in conjunction with option models with O_2 control	•	-	8612-0028
uick sample access	vewing window (13 x 25 cm) in the inner door for quick access	•	-	8612-0032

ACCESSORIES

Designation	Description	CB 260	*	ArtNo.
	for simple logging and documentation requirements with up to 5 networked units.			
APT-COM [™] 4 BASIC-Edition	version 4, BASIC edition	•	-	9053-0039
APT-COM [™] 4 GLP-Edition	for working under GLP-compliant conditions. Measured values are documented in a tamper-proof way in line with the requirements of FDA Regulation 21 CFR 11.			

Data sheet Model CB 260

BINDER Best conditions for your success

Designation	Description	CB 260	*	ArtNo.
	version 4, GLP edition	•	-	9053-0042
APT-COM™ 4 PROFESSIONAL- Edition	convenient unit and user management built on the BASIC edition. Suitable for networking up to 100 units.			
	version 4, PROFESSIONAL edition	•	-	9053-0040
Base	the base equipped with casters is used for the secure positioning and leveling of a \ensuremath{BINDER} CO2 incubator	•	-	9051-0044
	for CO2, consisting of a gas tank pressure regulator with connection parts and 5-meter hose	•	-	8012-0014
as cylinder connection set	for N2, consisting of a gas tank pressure regulator with connection parts and 5-meter hose	•	-	8012-0016
	for O_2 , consisting of a gas tank pressure regulator with connection parts and 5-meter hose	•	-	8012-0015
Gas tank changer	external, BINDER Gas Supply Service, for connecting 2 gas tanks (CO ₂ , N ₂ or O ₂), with audible and visual alarms, as well as zero-voltage alarm output	•	-	8012-2344
H-neutral detergent	concentrated, for gentle remove of residual contaminants; 1 kg	•	-	8012-2250
	IQ/OQ/PQ documents - supporting documents for validation performed by customers, according to customer requirements, PQ section added to qualification folder IQ/OQ; parameters: temperature, CO ₂ , O ₂ - or pressure, depending on unit			
	Digital in PDF format	•	-	7057-0005
	Hard copy inside folder	•	-	7007-0005
Qualification documents	IQ/OQ documents - supporting documents for validation performed by customers, consisting of: IQ/OQ checklists incl. calibration guide and comprehensive unit documentation; parameters: temperature, CO2, O2, pressure, depending on unit			
	Digital in PDF format	•	-	7057-0001
	- Hard copy inside folder	•	-	7007-0001
Rubber pads	set anti-slip feet	•	-	8012-0702
ihaker "Professional"	Shaker and controller equipped with LEMO connectors, exclusively for CB with electric access port, 100 - 240 V, 50/60 Hz $$	•		8012-1772
haker "Universal"	Shaker and controller connected via ribbon cable, universally applicable, 100 - 240 V, 50/60 Hz	•	-	8012-1771
	Stainless steel	•	-	8012-2183
helf, perforated	for divided inner door			
	- Stainless steel, divided	•	-	8012-1996
tacking adapter	for the thermally isolated stacking of two BINDER CO2 incubators	•	-	9051-0039
tacking frame	stable, vibration-free stacking frame on casters with stop brake for safe stacking of two CB series units	•	-	9051-0041

SERVICES



JLNVICLJ			
Designation	Description	*	ArtNo.
Calibration services			
Calibration certificate, temperature	Calibration of one (1) test temperature specified by the user in center of chamber, including certificate	-	DL30-0140
	Extension of calibration of one (1) additional test temperature specified by the user in the center of the usable space, including certificate	-	DL30-0102
CO2 calibration	CO_2 calibration at a specified value, measurements are taken with analyzed test gas at 5%, including certificate	-	DL30-0401
O2 calibration	O_2 calibration with 1 measuring point in center of chamber at a specified O_2 percentage value, including certificate	-	DL30-0402
Temperature measurement, 9 measuring points	Temperature measurement with 9 measuring points with a set value specified by the user, including certificate	-	DL30-0109
Temperature measurement, 18 measuring points	Temperature measurement with 18 measuring points with a set value specified by the user, including certificate	-	DL30-0118
Temperature measurement, 27 measuring points	Temperature measurement with 27 measuring points with a set value specified by the user, including certificate	-	DL30-0127
Installation services			
Unit commissioning	Connect the unit to the customer-side connections (electricity, water, wastewater, gas), basic functions check, brief operating instructions. (excl.: unpacking, setup, controller instructions, programming, installation work)	-	DL10-0110
Unit instructions	Instruction regarding operating principle and basic functions of the unit, operation of the control electronics including programming	-	DL10-0510
Maintenance contracts			
BRONZE 3-year maintenance contract	Maintenance service as contractually agreed, visual inspection of mechanical and electrical components, check of control response, 20% discount on spare parts	-	DL20-0710
GOLD 3-year maintenance contract	Maintenance service as contractually agreed, visual inspection of mechanical and electrical components, check of control response, 20% discount on spare parts, testing of all key functions, replacement of wear parts, calibration of one test temperature specified by the user in the center of the usable space, including certificate	-	DL20-0975
SILVER 3-year maintenance contract	Maintenance service as contractually agreed, visual inspection of mechanical and electrical components, check of control response, 20% discount on spare parts, testing of all key functions, calibration of one test temperature specified by the user in the center of the usable space, without certificate	_	DL20-0875
Maintenance services			
Maintenance	One-off maintenance service in accordance with maintenance schedule. Visual inspection of mechanical and electrical components, testing of all key functions. Calibration of a test temperature specified by the user in center of usable space without certificate	-	DL20-0608
Validation services			
Execution of IQ/OQ	Execution of IQ/OQ in accordance with qualification folder	-	DL41-0200
Execution of IQ/OQ/PQ	Execution of IQ/OQ/PQ in accordance with qualification folder	_	DL44-0500
Warranty service			
1-year warranty extension	The warranty is extended by 1 year from the delivery date, wear parts are excluded	-	DL50-0020

NOTES

- 01 Condensation may occur in the area around the access port. Access ports may be placed in custom locations for an additional charge.
- 02 UL mark is not granted when this option is used.

BINDER GmbH Tuttlingen, Germany TEL +49 7462 2005 0 info@binder-world.com www.binder-world.com

BINDER Asia Pacific (Hong Kong) Ltd. Kowloon, Hong Kong, P.R. China TEL +852 39070500 asia@binder-world.com www.binder-world.com BINDER Inc. Bohemia, NY, USA TEL +1 631 224 4340 usa@binder-world.com www.binder-world.us BINDER Environmental Testing Equipment (Shanghai) Co., Ltd. Shanghai, P.R. China TEL +86 21 685 808 25 china@binder-world.com www.binder-world.com