

## PRODUCT DATA SHEET

Standing: 2021-07-21

LAUDA Varioshake VS 150 OI  
Shaking incubator 230 V; 50/60 Hz  
Part Number: L003054

### Features

- Shaking incubator with orbital movement for precise and reproducible research
- Temperature display and setting digitally via LED display, in steps of 0.1 °C
- User-friendly microprocessor technology
- Membrane keyboard with distinctive symbols.
- PID type temperature control
- Limit values for over and under temperature alarm (optical and acoustic) can be set
- Frame for holding two trays
- Capacity 2 shaking trays; for vessels higher than 180 mm 1 tray, only
- Timer, microprocessor-controlled, with remaining time display and acoustic signal at the end of the process
- Timer adjustable from 1 min. up to 999 hours
- Interior lighting
- Interior cabinet, inside of doors and shaking table made of stainless steel
- Cooling coil for applications below room temperature
- Switching off of the shaking movement and the heating when the doors are opened
- Additional electromechanical temperature limiter
- RS 232 serial interface for trouble-free data transfer
- Drive by an overload-protected AC motor
- Compact, low-wear mechanism with mass balancing for a secure stand
- Constant speed, independent of load
- Excellent for continuous use
- Outer housing made of electrolytically galvanized sheet steel, powder-coated



Reserve technical changes



Working temperature min.  
28 °C



Working temperature max.  
70 °C

LAUDA DR. R. WOBSEY GMBH & CO. KG  
Pfarrstraße 41/43 • 97922 Lauda-Königshofen  
Postfach 1251 • 97912 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0 • F + 49 (0) 9343 503-222  
info@lauda.de • www.lauda.de  
WEEE-Reg-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen  
Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:  
LAUDA DR. R. WOBSEY Verwaltungs-GmbH  
Sitz Lauda-Königshofen  
Registergericht Mannheim • HRB 560226

Geschäftsführer:  
Dr. Gunther Wobser (Vors.), Dr. Mario Englert,  
Dr. Marc Stricker  
Beirat: Dr. Gerhard Wobser

# PRODUCT DATA SHEET

Standing: 2021-07-21

LAUDA Varioshake VS 150 OI  
 Shaking incubator 230 V; 50/60 Hz  
 Part Number: L003054

## Technical Features

Working temperature range	28 ... 70 °C
Working temperature range with external cooling	20 ... 70 °C
Ambient temperature range	10 ... 30 °C
Heater power max.	0.5 kW
Power consumption max.	0.8 kW
Shaking amplitude	25 mm
Shaking frequency	20 ... 250 1/min
Load max.	20 kg
Useable volume	157 L
Interior dimensions (WxDxH)	674 x 540 x 430 mm
Overall dimensions (WxDxH)	930 x 890 x 820 mm
Power plug	Power cord with angled plug (CEE7/7)
Power supply	230 V; 50/60 Hz

Reserve technical changes

LAUDA DR. R. WOBSEY GMBH & CO. KG  
 Pfarrstraße 41/43 • 97922 Lauda-Königshofen  
 Postfach 1251 • 97912 Lauda-Königshofen • DE

T + 49 (0) 9343 503-0 • F + 49 (0) 9343 503-222  
 info@lauda.de • www.lauda.de  
 WEEE-Reg-Nr.: DE 66 42 40 57

Kommanditgesellschaft: Sitz Lauda-Königshofen  
 Registergericht Mannheim • HRA 560069

Persönlich haftende Gesellschafterin:  
 LAUDA DR. R. WOBSEY Verwaltungs-GmbH  
 Sitz Lauda-Königshofen  
 Registergericht Mannheim • HRB 560226

Geschäftsführer:  
 Dr. Gunther Wobser (Vors.), Dr. Mario Englert,  
 Dr. Marc Stricker  
 Beirat: Dr. Gerhard Wobser